

EC DECLARATION OF CONFORMITY

MANUFACTURER

Rexxam Co.,Ltd. Kagawa factory
958, Ikeuchi, Konan-cho, Takamatsu, Kagawa 761-1494 JAPAN

AUTHORIZED EUROPEAN REPRESENTATIVE

REXXAM ELECTRONICS IRELAND CO.LIMITED
Donore Road, Drogheda, County Louth, Ireland

MEDICAL DEVICE

Model Name : Auto Lensmeter
Model number : Auto Lensmeter DL-900
Classification : Class I
Serial number : effective from 340011J

The undersigned hereby declares that the medical device as specified above conforms with the essential requirements listed in Annex I of EC Directive 93/42/EEC.

This declaration of conformity is based on the EC Directive 93/42/EEC Annex VII, with reference to articles 1 and 3 of directive 93/42/EEC.

The following standards are applicable:

EN 60601-1:60601-1 : 2006/AC:2010

Medical electrical equipment
Part 1:General requirements for safety

EN 60601-1-2:2007:AC:2010

Medical electrical equipment
Part 1-2:General requirements for safety
Collateral standard : Electromagnetic compatibility -
Requirements and tests

Date CE Mark was affixed : April. 24, 2014

Kagawa Factory, June.24, 2014

Place and date of issue



Name, signature and
position of manufacture

Kagawa Factory Manager of Development Dept.1

Masakatsu Iwamoto