

DECLARATION OF CONFORMITY

MANUFACTURER

Rexxam Co.,Ltd. Kagawa factory
958 Ikeuchi, Konan-cho, Takamatsu-city, Kagawa 761-1494, Japan

AUTHORIZED EUROPEAN REPRESENTATIVE

Rexxam Czech s.r.o.
Prumyslova 937, 334 01 Prestice, Czech Republic

MEDICAL DEVICE

Model Name : REF TOPO RET-700
Classification : Class I (Devices with a measuring function)
Serial number : effective from 97DD0011(Valid until a product is changed.)

This declaration is only valid for the manufactured products with the document released by us of each product.

The undersigned hereby declares that the medical device as specified above conforms with the essential requirements listed in Annex I of EC Directive 93/42/EEC.

This declaration of conformity is based on the EC Directive 93/42/EEC Annex V,VI and supported by the TUV Rheinland LGA Products GmbH(0197) (TILLYSTRASSE 2, 90431 NUREMBERG · GERMANY·) Annex V certificate, with reference to articles 3 of directive 93/42/EEC.

We declare that we are exclusively responsible for this declaration.

Kagawa Factory, 2.November.2017

Place and date of issue



Name, signature and
position of manufacture
Development Dept.1 Manager
Shunji Takashima